



## Static Trial

During the static trial, the subject should:

1. Stand stationary 2. Feet on one or two of the force platforms 3. Comfortable stance 4. Feet shoulder width apart 5. Arms out to so that markers are not occluded During the static trial, the user should:

1. Instruct the subject where to stand and in what posture 2. Collect approximately a three second trial using the motion capture software while the subject is stationary 3. Instruct the subject to relax 4. Label the static trial to ensure all targets appear If all targets have been placed on the subject, if not, collect a new static trial, or overwrite the original static trial

## Dynamic Gait Trials

Collect dynamic trials:

1. Have the patient start at one end of the capture volume and have them walk across the force plates at a self selected normal speed. 2. Collect at least one good foot strike per gait trial. [Click here to find out more information about a "good" foot strike](#) 2. Make sure: 1. No targets fall off the subject during data collection 2. Targets do not become occluded due to clothing, etc. To complete the IORgait analysis, at least two clean foot strikes (one right, one left) must be collected. However, the total number of trials collected is an individual lab decision. Many users collect at least five clean foot strikes for the left and five for the right. The amount of trials this requires may depend on the number of force platforms in your lab.

From:  
<https://has-motion.com/wiki/> - Wiki Documentation Site

Permanent link:  
[https://has-motion.com/wiki/doku.php?id=other:ior\\_gait:documentation:gait\\_collection&rev=1718717126](https://has-motion.com/wiki/doku.php?id=other:ior_gait:documentation:gait_collection&rev=1718717126)

Last update: 2024/06/18 13:25

